

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 07 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130142 ✓
Name		Scott E. Gardner	ID# 6222 Date 30 Apr 09 ✓
A	Agency	Directorate of Emergency Services	Phone # 384-3888
Instrument Location		Bldg #656, A St, Ft Richardson, AK 99505 ✓	
Alco S/N		78820 ✓	Target Value .079 ✓ High Pressure 900 psi
B	Alco Test Values	1 st Alco .085 ✓	2 nd Alco .085 ✓
Signature		✓	
		COB 5/13/09	
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of May 2009.

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130142

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130142

APRIL 30, 2009

OPERATOR'S NAME:

BARNER/SCOTT/E

OPERATOR'S NUMBER: 6222

SUBJECT'S LAST NAME:

VERIFICATION OF CALIBRATION

SUBJECT'S FIRST NAME/MI :

N/A

O.L. #: N/A

DEPT/AGENCY: FRM4

CASE/REPORT: N/A

TEST TYPE: V

ALCO TARGET VALUE: .079

ALCO S/N: 72829

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.27 in
ALCO TARGET .079 11:36
BLANK TEST .000 11:38
INTERNAL STANDARD VERIFIED 11:38
ALCO TV 30.29 in .085 11:38
BLANK TEST .000 11:39
SUBJECT SAMPLE .000 11:39
BLANK TEST .000 11:40
ALCO TV 30.29 in .085 11:41
BLANK TEST .000 11:41

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130142

APRIL 30, 2009
TIME 11:42

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 03/29/00

HEATERS

SAMPLE CHAMBER: 43c

BREATH TUBE: 43c

BAROMETER: 30.29 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNOPQRSTUVWXYZ[\] ^ _ `abdefghijklmnop
qrstuvwxyz{|}~